

ENP Target Hall D Shutdown Checklist

Document ID: 999

Group: ENP Target

Subsystem: Hall D Shutdown

Author: Smith, Elton (elton)

Submitted By: Smith, Elton (elton)

Revision Number: 1

Revision Date: 11-Sep-2018 16:55

Revision Comment: Checklist based on shutdown before Hurricane Florence

1. Ensure that both H2 tanks are valved off
2. Turn off the turbo station that is pumping on the OVC
3. Turn off the pulse tube compressor and kill power at the service disconnect